

ENGRX CORP
4638 90 A AVENUE
EDMONTON AB T6B 2P9

Date: July 9, 2015
Account #: 58869
Journal #: 62981
Our File #: 5544774

Attn: RICK GERELUK

Re: Application for Design Registration

The design, as detailed in your, 1 NPS Flowmeter, for a Fitting is accepted for registration as follows:

Registered To: FOX THERMAL INSTRUMENTS **CRN:** 0F6270.1

MDMT: -20 deg F

MAWT: 100 deg F

MAWP: 500 psig

Drawing #: 100381, 105288, 100728

Drawing Revision: F / A / K(as noted)

Conditions Of Registration:

- 1- Fitting Registration Expiry Date; July 08, 2025.
- 2- The registration is valid until the indicated expiry date only if the Manufacturer maintains a valid quality management system approved by an acceptable third-party agency until that date. Should the approval of the quality management system lapse before the expiry date indicated above, this registration shall become void.

Contact me if you have any questions. The invoice for registration will be forwarded under separate cover.

ALIREZA BONAKDAR
(778) 396-2032
Alireza.Bonakdar@safetyauthority.ca
Design Engineer

cc: