



EC Declaration of Conformity

No. F-157

Fox Thermal Instruments

399 Reservation Road

Marina CA, 93933

We hereby declare that the equipment described below complies fully with the requirements of the Pressure Equipment Directive 97/23/EC and 2014/68/EU Article 13.

Description: Thermal Mass Flow Meter – Model FT2A

Conformity Assessment Modules followed – D1

Name and address of the Notified Body monitoring the manufacturer's quality system: -

BSI, Kitemark Court, Davy Avenue, Knowlhill, Milton Keynes, MK5 8PP

Notified Body number - 0086

Harmonized Standards applied: - EN ISO 15614-1 and EN ISO 9606-1

Other technical standards used: - ASME B31.3

Other Community Directives applied: -

2014/30/EU EMC Directive; Emissions and Immunity Testing: EN61326-1:2013

2014/35/EU Low Voltage Directive (LVD); Product Safety Testing: EN61010-1:2010

I hereby declare that the Fox Thermal Instruments FT2A has been designed to comply with the relevant sections of the above referenced specifications. The FT2A complies with all applicable Essential Requirements of the Directives.

Signed:



Name: Brad Lesko Title: President/CEO

Place Marina, California Date 4/8/16