



CE Declaration of Conformity

No. F-168 Rev. A
Fox Thermal Instruments
399 Reservation Road
Marina CA, 93933

Fox Thermal Instruments declares that the equipment described below complies fully with the requirements of CE.

Description: Thermal Mass Flow Meter- Model FT4X

This equipment has been tested against the following harmonized standards:

2014/30/EU	EMC (Electromagnetic Compatibility Directive): Electrical Equipment for Measurement, Control and Laboratory Use EN61326-1:2013
94/9/EC & 2014/34/EU	ATEX Directive; Product Safety Testing: EN 60079-0:2012 +A11:2013, EN 60079-1:2014, EN 60079-31:2014, EN 60529:1991 +A1:2000 IECEX scheme: IEC 60079-0:2011, IEC 60079-1:2007, IEC 60079-31:2008, IEC 60529:2001

I hereby declare that the Fox Thermal Instruments FT4X has been designed to comply with the relevant sections of the above referenced specifications. The FT4X complies with all applicable Essential Requirements of the Directive.

Signed:

 1/15/2018

Name: Brad Lesko Title: President/ CEO

Place: Marina, California Date: 1/15/2018