



399 Reservation Rd
Marina, CA 93933, USA
PH: (831) 384-4300 FAX: (831) 384-4312
Website: www.foxthermalinstruments.com

Zero CAL-CHECK™ CERTIFICATE

CALIBRATION VALIDATION

Zero CAL-CHECK™ Performed on: 9/13/12 3:04:46 PM

Firmware version: v3.65

Fox Meter Serial Number: 16505

Type of Test: Out of Pipe Using the Factory Baseline

Zero CAL-CHECK™ Results: PASS

Zero CAL-CHECK™ Value (Volts): 0.27050

Zero CAL-CHECK™ Error (Volts): 0.00253

Zero CAL-CHECK™ Error (%): 0.94 %

Test Temperature 85.2°F

Tag #/Meter Location: Compressor #528

Test performed by: John Doe

Additional Comments: _____

Zero CAL-CHECK™ is a calibration routine that validates the flow meter's calibration accuracy by testing the following:

- * Sensor free of residue or film
- * Zero Stability of the meter
- * Thermal Conductivity (heat transfer) repeatability of the sensor

At the conclusion of the test, the meter will display a pass/fail message and the Zero CAL-CHECK™ data.

A "pass" result confirms the meter is measuring accurately.