



399 Reservation Road
Marina, CA 93933 USA
Phone: 831-384-4300
Fax: 831-384-4312
sales@foxthermal.com

Zero CAL-CHECK[®] CERTIFICATE CALIBRATION VALIDATION

Zero CAL-CHECK[™] Performed on: September 17 2020 10:53:41 AM

Firmware version: FT3 V3.13

Fox Meter Serial Number: 16045

Zero CAL-CHECK[™] Results: PASS

Zero CAL-CHECK[™]: -1.51%

Test Temperature 78.6 F

Tag #/Meter Location: Tag #3

Test performed by: Patrick

Additional Comments: App ID 540. Test meter #64

Zero CAL-CHECK[®] is a calibration routine that validates the flow meter's calibration accuracy by testing the following:

- * Sensor free of residue or film
- * Zero Stability of the meter
- * Thermal Conductivity (heat transfer) repeatability of the sensor

At the conclusion of the test, the meter will display a pass/fail message and the Zero CAL-CHECK[®] data.

A "pass" result confirms the meter is measuring accurately.