



399 Reservation Road
Marina, CA 93933 USA
Phone: 831-384-4300
Fax: 831-384-4312
sales@foxthermal.com

FT4X CAL-V™ CERTIFICATE

CALIBRATION VALIDATION

CAL-V™ Performed on: September 16 2020

7:49:53 AM

Firmware version: FT4X V7.0

Fox Meter Serial Number: F07455

CAL-V™ Results: Pass

CAL-V™: 0.11

Test Temperature 73.6 F

Tag #/Meter Location: Tag #1

Test performed by: Patrick

Additional Comments: App ID 10613. Test meter #42

CAL-V™ is a calibration routine that validates the flow meter's calibration accuracy by testing the following:

- * Repeatability of sensor
- * Repeatability of sensor electronics
- * Confirms Calibration Algorithms

At the conclusion of the test, the meter will display a pass/fail message and the CAL-V™ data.

A "pass" result confirms the meter is measuring accurately.

CAL-V™ limits: $\pm 0 - 0.8$ Pass, $\pm 0.8-1.0$ Warning, $> \pm 1.0$ Fail

Configuration:

Pipe Diameter:	127 mm	Gas Select:	Single Gas
Customer STP:	60.0 Deg F & 14.73 Psia	CO2	100%
4-20 mA Range:	0 - 3793 KG/Hr		
Zero Flow Cutoff:	11.094147 KG/Hr		
Previous CAL-V:	0.19		
Previous CAL-V:	Pass		
Gross Heating Value[BTU/FT3]:	0		
Density[Kg/M3]:	1.8725		